BIOSYSTEMS A25

URIC ACID TBHBA

Enzymatic, Colorimetric

Liquid
2 Reagents

REF		Cont.	
D95459	**5 x 100 ml**	4 x 100 ml	Reagent 1
		1 x 100 ml	Reagent 2
D98461	**5 x 50 ml**	4 x 50 ml	Reagent 1
		1 x 50 ml	Reagent 2

D94708	1 x 3 ml	Uric acid Standard	
D98485	5 x 3 ml	Calibrator	Diacal Auto
D98481	12 x 5 ml	Control normal	Diacon N
D98482	12 x 5 ml	Control abnormal	Diacon P

1. Reagent preparation:

The reagents are ready to use. For Sample Start mix 4 parts R1 with 1 part R2.

2. Instrument settings:

Temperature: 37 °C

GENERAL	Test name	Uric Acid TBHBA
	Analysis mode	Endpoint
	Sample type	Serum
	Units	mg/dL
	Reaction type	Ascending
	Decimals	1
	Replicates	1
	Name of assoc. Constituent	-
PROCEDURE	Type of reading	Monochromatic
Volumes	Sample	4
	Reagent 1	200
	Reagent 2	50
	Washing	1.2
	Predilution factor	-
	Postdilution factor	2
Filters	Main	505
	Reference	-
Times	Reading 1	600
	Reading 2	-
	Reagent 2	300
CALIBRATION	Type of calibration	Multiple Calibrator
	Calibrator replicates	3
	Blank replicates	1
	Calibration curve	-
OPTIONS	Blank absorbance limit	0,180
	Kinetic blank limit	-
	Linearity limit	20

NOTE:
These suggested instructions and instrument parameters are to be used in conjunction with the reagent package insert and the instrument operation manual. Refer to these documents for complete instructions before performing the tests.